


Lampiran 7. Hasil Uji Timbal



UNIVERSITAS ISLAM INDONESIA LABORATORIUM TERPADU

LAB. INSTRUMENTASI, FISIKA DASAR DAN KIMIA DASAR
 Jl Kaliurang Km 14,5 Yogyakarta 55584 Telp. (0274)895920 ext. 3045, 3016, Fax (0274) 896439 ext. 3020
 Website: <http://lab.uii.ac.id> , e-mail : lab.terpadu@uui.ac.id

No. Dok : Form-06/Hasil Uji Rev. 2
 Tgl. Terbit : 14 Maret 2013

Hasil Analisis Spektrofotometer Serapan Atom (SSA)

Sampel : Cair
 Kode Sampel : 1573/P/AAS
 Asal Sampel : Farmasi Ull
 Tanggal diterima : 8 Maret 2018
 Tanggal dianalisis : 9 Maret 2018
 Parameter : Pb

No	Sample ID	Seq No.	El	Standar	Mean Sig (Absorbance)	Limit Detection from standard	Mean Samp	Std Dev	Samp Units
1	Calib Blank	2	Pb	0	-0.0004			0.00020	mg/L
2	std 1	3	Pb	0.5	0.0018	0.0007		0.00020	mg/L
3	std 2	4	Pb	1	0.0048	0.0007		0.00030	mg/L
4	std 3	5	Pb	2	0.0099	0.0007		0.00030	mg/L
5	std 4	6	Pb	3	0.0152	0.0007		0.00010	mg/L
6	std 5	7	Pb	5	0.0230	0.0007		0.00020	mg/L
7	std 6	8	Pb	10	0.0454	0.0007		0.00080	mg/L
8									
9									
10	1570-1	9	Pb		0.0023		0.4525	0.00020	mg/L

Penyelia,

Gani Purwiantono, M.Sc

Jogjakarta, 9 Maret 2018
 Analisis,

 Yusuf Habibi, S.Si

0.0231 mg/g